




Wisconsin State
Laboratory of Hygiene

UNIVERSITY OF WISCONSIN-MADISON



A lesson in the value of thorough QC/QA as it applies to whole genome sequencing

Sara Wagner

Wisconsin State Laboratory of Hygiene

03/07/2019



Background

- Started routinely using MagNA pure LC for extraction Spring 2018
- Including a negative control on each extraction run.
 - Nanodrop
 - Qubit
- By mid August we implemented a bioinformatics pipeline to evaluate the quality of a sequenced isolate.
 - FastQC
 - Serotype
 - Contigs



Problem

- Specimens were exhibiting a higher than expected number of contigs (255-3173)
- Size range varied greatly between 2.7Mb-5.8Mb.
- Negative controls showed no signs of contamination through Qubit or Nanodrop



MultiQC

v1.6

General Stats

QUAST

Assembly Statistics

Number of Contigs

Trimmomatic

FastQC

Sequence Counts

Sequence Quality Histograms

Per Sequence Quality Scores

Per Base Sequence Content

Per Sequence GC Content

Per Base N Content

Sequence Length Distribution

Sequence Duplication Levels


Overrepresented sequences

Adapter Content

Number of Contigs

This plot shows the number of contigs found for each assembly, broken down by length.

Counts Percentages





Problem

- End of July 2018 – early August 2018 we had 8 runs with contaminated specimens.
- All specimens were extracted by MagNA pure LC
- No contamination in hand extracted specimens.



Troubleshooting

- Wipe tests were performed on the MagNa Pure
 - 16s – positive
 - Stx1 and Stx2 – positive
- Cleaning was performed on the machine and wipe tests were negative.
- A checkerboard extraction was performed
- Water tested with PCR - positive



- More cleaning was performed
 - When cleaning on the inside of the nozzle head where the pipettes are held, some black gunk was found and removed
- Another wipe test and checkerboard extraction was performed – Negative
- MagNa Pure extracted samples were library prepped and ran





Sources of contamination

- Hypothesis
 - Too much bug was being put into the machine for extraction?
 - Incorrect use of the machine?
 - Wells empty but chosen on the machine?
 - Insufficient decontamination between runs?



Future Steps

- Negative control taken through library preparation and sequencing
 - Specimens are library prepped and ran with the negative control that they were extracted with
- Implemented deeper cleaning of MagNa Pure on a weekly basis.
- Monthly wipe tests
- Continue to look at more metrics for our specimens to detect anomalies in real time.